

K3603S6
K3607
K3608

Information about CE labelling

The calibration units K36xx carry the CE label and fulfil the requirements of EU guideline 89/336/EEG "Electromagnetic compatibility" and its listed harmonised European standards (EN).

The EU declarations of conformity according to the above mentioned EU guideline, Article 10, for the relevant authorities are available from:

HOTTINGER BALDWIN MESSTECHNIK GMBH
Im Tiefen See 45
D-64 293 Darmstadt

The calibration units are designed for use in living areas and fulfil the requirements for interference immunity according to EN 50082-1:1992 and for emission according to EN 55011:1991 and EN 55022:1994.

In addition to the details given in the operating instructions, the following information should be followed when operating the device:

The maximum deviation of the measurement under the influence of electromagnetic fields of 3V/m in the frequency range 27 MHz to 500 MHz is 0.01% of the nominal value assuming the following requirements.

- K3603S6, K3607: The cable to the amplifier is connected to ST1; BU1...BU5 should not be used.
- All cables connected to the DK38 must be screened and the screen should be connected flat. This is provided with HBM Greenline cables.